




**FEHILY  
TIMONEY**

CONSULTANTS IN ENGINEERING,  
ENVIRONMENTAL SCIENCE  
& PLANNING

## **APPENDIX 7.2**

Equipment Calibration  
Certificates





# NSAI

## National Metrology Laboratory

### Certificate of Calibration

Issued to Fehily Timoney & Company  
J5 Plaza  
North Business Park  
North Road  
Dublin 11

Attention of John Mahon

---

Certificate Number	180300
Item Calibrated	Svantek SVAN 977 Sound Level Meter with ACO 7052E Microphone
Serial Number	34876 (SLM) and 56429 (Microphone)
Client ID Number	#2
Order Number	6252
Date Received	24 Jan 2018
NML Procedure Number	AP-NM-09

**Method** The above sound level meter was allowed to stabilise for a suitable period in laboratory conditions. It was then calibrated by carrying out the verification tests detailed in IEC 61672-3 (2006), *Periodic tests, specification for the verification of sound level meters*. This standard specifies a procedure for the periodic verification of conformance of a sound level meter or integrating-averaging meter to IEC 61672-1 (2003).

**Calibration Standards** Norsonic 1504A Calibration System incorporating:  
SR DS360 Signal Generator, No. 0735 [Cal Due Date: 21 Dec 2018]  
Agilent 34401A Digital Multimeter, No. 0736 [Cal Due Date: 17 Nov 2018]  
B&K 4134 Measuring Microphone, No. 0743 [Cal Due Date: 28 Apr 2019]  
B&K 4228 Pistonphone, No. 0740 [Cal Due Date: 21 Mar 2019]  
B&K 4226 Acoustical Calibrator, No. 0150 [Cal Due Date: 15 May 2018]

---

Calibrated by

  
David Fleming

Approved by

  
Paul Hetherington

Date of Calibration

31 Jan 2018

Date of Issue

31 Jan 2018



This certificate is consistent with Calibration and Measurement Capabilities (CMC's) that are included in Appendix C of the Mutual Recognition Arrangement (MRA) drawn up by the International Committee for Weights and Measures. Under the MRA, all participating institutes recognize the validity of each other's calibration certificates and measurement reports for quantities, ranges and measurement uncertainties specified in Appendix C (for details see [www.bipm.org](http://www.bipm.org))

**Issued to:**

**Fehily Timoney**  
J5 Plaza  
North Park Business Park  
North Road  
Dublin 11

**Calibration Reference**

SLM200095

**Test Date:** 03/06/2020

**Procedure:** TP-SLM-1

---

**Equipment**

<b>Item Calibrated:</b>	Sound Level Meter	<b>Model</b>	977
<b>Make:</b>	Svantek	<b>Serial Number:</b>	69552

---

**Calibration Procedure**

The sound level meter was allowed to stabilize for a suitable period, as described in the manufacturer's instruction manual, in laboratory conditions. The sound level meter was calibrated by carrying out the verification tests detailed in IEC 61672-3 (2006), Periodic tests, specification of sound level meters. Tolerances for verification procedures are specified in IEC 61672-1 (2003).

---

**Calibration Standards**

<b>Description</b>	<b>Serial Number</b>
National Instruments PXI-4461	19C91D2
Stanford Research DS360	123803

---

The standards used in this calibration are traceable to NIST and/or other National Measurement Institutes (NMI's) that are signatories of the International Committee of Weights and Measures (CIPM) mutual recognition agreement (MRA).

---

**Signed on behalf of Sonitus Systems:**



---



Self-generated noise - IEC 61672-3 Test #10

SLM Measuring Mode: Leq

SLM Configuration	Freq. Weighting Network	SLM Reading
Microphone Installed	A	18.6
Microphone replaced by electrical input device fitted with short circuit	A	7.3
	C	7.3
	Z	12.6

Acoustical signal test of a frequency weighting - IEC 61672-3 Test #11

Range: reference level range

Frequency Weighting: C

Time Weighting: Slow

Input	Freq	Expected Level	Deviation	Tol +/-
94 dB	1000 Hz	94.0	0.0	1.0
	125 Hz	93.7	0.2	1.0
	4000 Hz	92.3	0.5	1.0

The frequency response was tested using an electrostatic actuator. Appropriate correction factors were applied where available from the manufacturer's instruction manual.

Electrical tests of frequency weighting - IEC 61672-3 Test #12

Range: reference level range

A-weighting

Freq	Expected Level	SLM Reading	Deviation	Tol +	Tol -
63	95.0	95.1	0.1	1.5	-1.5
125	95.0	95.0	0.0	1.5	-1.5
250	95.0	94.9	-0.1	1.4	-1.4
500	95.0	95.0	0.0	1.4	-1.4
1000	95.0	95.0	0.0	1.1	-1.1
2000	95.0	95.1	0.1	1.6	-1.6
4000	95.0	95.1	0.1	1.6	-1.6
8000	95.0	95.2	0.2	2.1	-3.1
16000	95.0	94.8	-0.2	3.5	-17.0

C-weighting

Freq	Expected Level	SLM Reading	Deviation	Tol +	Tol -
63	95.0	95.0	0.0	1.5	-1.5
125	95.0	95.4	0.4	1.5	-1.5
250	95.0	95.0	0.0	1.4	-1.4
500	95.0	95.0	0.0	1.4	-1.4
1000	95.0	95.0	0.0	1.1	-1.1
2000	95.0	95.1	0.1	1.6	-1.6
4000	95.0	95.1	0.1	1.6	-1.6
8000	95.0	95.2	0.2	2.1	-3.1
16000	95.0	94.7	-0.3	3.5	-17.0

Linear

Freq	Expected Level	SLM Reading	Deviation	Tol +	Tol -
63	95.0	95.1	0.1	1.5	-1.5
125	95.0	95.1	0.1	1.5	-1.5
250	95.0	95.0	0.0	1.4	-1.4
500	95.0	95.0	0.0	1.4	-1.4
1000	95.0	95.0	0.0	1.1	-1.1
2000	95.0	95.0	0.0	1.6	-1.6
4000	95.0	95.1	0.1	1.6	-1.6
8000	95.0	95.1	0.1	2.1	-3.1
16000	95.0	95.1	0.1	3.5	-17.0

Frequency and Time Weightings at 1 kHz IEC 61672-3 Test #13

Range: reference level range

Time Weighting	Freq. Weighting	Expected Level	Deviation	Tol +/-
Fast	A	94.0	ref	
	C	94.0	0.0	0.2
Slow	A	94.0	0.0	0.2
LEQ	A	94.0	0.0	0.2

Linearity level on reference range - IEC 61672-3 Test #14  
 Input frequency: 8 kHz  
 SLM Measuring Mode: SPL

Range	Expected Level	SLM Reading	Deviation	Tol +/-
123 dB	94.0	94.0	0.0	1.1
	99.0	99.0	0.0	1.1
	104.0	104.0	0.0	1.1
	109.0	109.0	0.0	1.1
	114.0	114.0	0.0	1.1
	119.0	119.0	0.0	1.1
	124.0	124.0	0.0	1.1
	129.0	129.0	0.0	1.1
	134.0	134.0	0.0	1.1
	135.0	135.0	0.0	1.1
	136.0	136.0	0.0	1.1
	137.0	137.0	0.0	1.1
	89.0	89.0	0.0	1.1
	84.0	84.0	0.0	1.1
	79.0	79.0	0.0	1.1
	74.0	74.0	0.0	1.1
	69.0	69.0	0.0	1.1
	64.0	64.0	0.0	1.1
	59.0	59.0	0.0	1.1
	54.0	54.0	0.0	1.1
	49.0	49.0	0.0	1.1
	44.0	44.0	0.0	1.1
	39.0	39.0	0.0	1.1
	38.0	38.0	0.0	1.1
	37.0	37.0	0.0	1.1
	36.0	36.0	0.0	1.1
	35.0	35.0	0.0	1.1

Toneburst response - IEC 61672-3 Test #16

Range: reference level range

Burst Type	Response	Expected Level	SLM Reading	Deviation	Tol +	Tol -
0.25 ms	LAF <sub>MAX</sub>	111.0	110.9	-0.1	0.8	-0.8
2.0 ms	LAF <sub>MAX</sub>	120.0	119.9	-0.1	1.3	-1.3
200 ms	LAF <sub>MAX</sub>	137.0	137.0	0.0	1.3	-3.3
2.0 ms	LAS <sub>MAX</sub>	111.0	111.3	0.3	0.8	-0.8
200 ms	LAS <sub>MAX</sub>	130.6	130.6	0.0	1.3	-3.3

Peak C sound level - IEC 61672-3 Test #17

Range: reference level range

Pulse Type	Freq	Expected Level	SLM Reading	Deviation	Tol +/-
1 cycle	8 kHz	135.4	135.3	-0.1	2.4
Pos ½ cycle	500 Hz	137.4	137.3	-0.1	1.4
Neg ½ cycle	500 Hz	137.4	137.3	-0.1	1.4

Overload indication IEC 61672-3 Test #18

Test Description	Overload at	Meas. Diff. (Pos – Neg)	Tol +/-
Pos. ½ cycle at 4 kHz	140.0		
Neg. ½ cycle at 4 kHz	140.0		
Level difference		0.0	1.8

**Calibration Notes**

1. The manufacturer's instruction manual was accessed through the manufacturer's website.
2. The sound level meter was powered by a regulated 9V power supply provided by the testing laboratory.



# NSAI

## National Metrology Laboratory

### Certificate of Calibration

Issued to Fehily Timoney & Company  
J5 Plaza  
North Park Business Park  
North Road  
Dublin 11

Attention of John Mahon

---

Certificate Number	173660
Item Calibrated	Svantek SVAN 977 Sound Level Meter with ACO 7052E Microphone
Serial Number	34173 (SLM) and 54691 (Microphone)
Client ID Number	#3
Order Number	6308
Date Received	20 Oct 2017
NML Procedure Number	AP-NM-09

**Method** The above sound level meter was allowed to stabilise for a suitable period in laboratory conditions. It was then calibrated by carrying out the verification tests detailed in IEC 61672-3 (2006), *Periodic tests, specification for the verification of sound level meters*. This standard specifies a procedure for the periodic verification of conformance of a sound level meter or integrating-averaging meter to IEC 61672-1 (2003).

**Calibration Standards** Norsonic 1504A Calibration System incorporating:  
SR DS360 Signal Generator, No. 0735 [Cal Due Date: 18 Nov 2017]  
Agilent 34401A Digital Multimeter, No. 0736 [Cal Due Date: 07 Nov 2017]  
B&K 4134 Measuring Microphone, No. 0742 [Cal Due Date: 18 Jan 2018]  
B&K 4228 Pistonphone, No. 0741 [Cal Due Date: 08 Jan 2018]  
B&K 4226 Acoustical Calibrator, No. 0150 [Cal Due Date: 15 May 2018]

---

Calibrated by



David Fleming

Approved by



Paul Hetherington

Date of Calibration

31 Oct 2017

Date of Issue

01 Nov 2017



This certificate is consistent with Calibration and Measurement Capabilities (CMC's) that are included in Appendix C of the Mutual Recognition Arrangement (MRA) drawn up by the International Committee for Weights and Measures. Under the MRA, all participating institutes recognize the validity of each other's calibration certificates and measurement reports for quantities, ranges and measurement uncertainties specified in Appendix C (for details see [www.bipm.org](http://www.bipm.org))

ENVIRONMENTAL CONDITIONS

Temperature	Relative humidity	Ambient pressure
20 °C	37%	1005 hPa

TEST EQUIPMENT

Item	Manufacturer	Model	Serial no.	Description
1.	SVANTEK	SVAN 401	127	Signal generator
2.	SVANTEK	SVAN 912A	9537	Sound & Vibration Analyser
3.	KEITHLEY	2000	0910165	Digital multimeter
4.	SVANTEK	SV33	48878	Acoustic calibrator
5.	SVANTEK	ST00	-	Microphone equivalent electrical impedance (18pF)
6.	DYTRAN	3335A	1376	Reference accelerometer

CONFORMITY & TEST DECLARATION

1. Herewith Svantek company declares that this instrument has been calibrated and tested in compliance with the internal ISO9001 procedures and meets all specification given in the Manual(s) or respectively surpass them.
2. The acoustic calibration was performed using the Sound Calibrator and is traceable to the GUM (Central Office of Measures) reference standard - sound level calibrator type 4231 No 2292773.
3. The vibrational calibration was performed using the Back-to-Back Comparison method and is traceable to the GUM (Central Office of Measures) reference standard - accelerometer type 8305 No 1435233.
4. The information appearing on this sheet has been compiled specifically for this instrument. This form is produced with advanced equipment & procedures which permit comprehensive quality assurance verification of all data supplied herein.
5. This calibration sheet shall not be reproduced except in full, without written permission of the SVANTEK Ltd.

Calibration specialist: Krzysztof Kubel



Test date: 2018-06-01



ISO9001 certified

FACTORY CALIBRATION DATA OF THE SVAN 977A No. 69557

with preamplifier SVANTEK type SV12L No. 72140 and microphone ACO PACIFIC type 7052E No. 69617

SOUND LEVEL METER

1. CALIBRATION (electrical)

LEVEL METER function; Characteristic: A;  $f_{ref}$ =1 kHz; Input signal =110.9 dB;

Range	Low (120dB)	High (137dB)
Indication [dB]	114.0	114.0
Error [dB]	0.0	0.0

2. CALIBRATION\* (acoustical)

LEVEL METER function; Range: High; Reference frequency: 1000 Hz; Sound Pressure Level: 113.99 dB.

Characteristic	Correct value [dB]	Indication [dB]	Error [dB]
Z	113.99	113.97	-0.02
A	113.99	113.97	-0.02
C	113.99	113.97	-0.02

Calibration measured with the microphone ACO PACIFIC type 7052E No. 69617. Calibration factor: 0.02 dB.

3. LINEARITY TEST\* (electrical)

LEVEL METER function; Range: Low; Characteristic: A;  $f_{ref}$ = 31.5 Hz

Nominal result LEQ [dB]	24.0	25.0	26.0	28.0	30.0	40.0	60.0	80.0
Error [dB]	0.1	0.1	0.0	0.0	0.0	0.0	0.0	0.0

LEVEL METER function; Range: Low; Characteristic: A;  $f_{ref}$ = 1000 Hz

Nominal result LEQ [dB]	24.0	25.0	26.0	28.0	30.0	40.0	60.0	80.0	100.0	120.0
Error [dB]	0.1	0.0	0.0	-0.0	-0.0	-0.0	-0.0	-0.0	0.0	-0.0

LEVEL METER function; Range: Low; Characteristic: A;  $f_{ref}$ = 8000 Hz

Nominal result LEQ [dB]	24.0	25.0	26.0	28.0	30.0	40.0	60.0	80.0	100.0	119.0
Error [dB]	0.0	-0.0	-0.0	-0.0	-0.0	-0.0	-0.0	-0.0	-0.0	-0.0

LEVEL METER function; Range: High; Characteristic: A;  $f_{ref}$ = 31.5 Hz

Nominal result LEQ [dB]	35.0	36.0	37.0	38.0	40.0	60.0	80.0	97.0
Error [dB]	0.1	0.1	0.1	0.0	0.0	-0.0	-0.0	0.0

LEVEL METER function; Range: High; Characteristic: A;  $f_{ref}$ = 1000 Hz

Nominal result LEQ [dB]	35.0	36.0	37.0	38.0	40.0	60.0	80.0	100.0	120.0	137.0
Error [dB]	0.1	0.1	0.0	0.0	0.0	-0.0	-0.0	0.0	-0.0	-0.0

LEVEL METER function; Range: High; Characteristic: A;  $f_{ref}$ = 8000 Hz

Nominal result LEQ [dB]	35.0	36.0	37.0	38.0	40.0	60.0	80.0	100.0	120.0	136.0
Error [dB]	0.0	0.0	0.0	-0.0	0.0	-0.0	-0.0	-0.0	-0.0	0.0

4. TONE BURST RESPONSE\*

LEVEL METER function; Characteristic: A;  $f_{ref}$ = 4000 Hz; Burst duration: 2 s

Range: Low; Steady level nominal result = 117dB

Result	Detector	Duration [ms]	1000	500	200	100	50	20	10	5	2	1	0.5	0.25
MAX	Fast	Indication [dB]	117.0	116.9	116.0	114.4	112.2	109.7	105.8	102.9	99.9	95.9	92.9	89.9
		Error [dB]	0.0	0.0	0.0	0.0	-0.0	-0.0	-0.1	0.0	-0.0	-0.1	-0.1	-0.1
	Slow	Indication [dB]	115.0	112.9	109.6	106.8	103.9	100.0	97.0	94.0	90.0	-	-	-
		Error [dB]	0.0	0.1	-0.0	-0.0	-0.0	-0.0	-0.0	-0.0	-0.0	-	-	-
SEL	-	Indication [dB]	117.0	114.0	110.0	107.0	104.0	100.0	97.0	94.0	90.0	86.9	83.9	80.9
		Error [dB]	0.0	-0.0	-0.0	0.0	-0.0	0.0	0.0	-0.0	-0.0	-0.0	-0.1	-0.1

**Issued to:**

**Fehily Timoney**  
J5 Plaza  
North Park Business Park  
North Road  
Dublin 11

**Calibration Reference**

SLM200094

**Test Date:** 03/06/2020

**Procedure:** TP-SLM-1

---

**Equipment**

<b>Item Calibrated:</b>	Sound Level Meter	<b>Model</b>	977
<b>Make:</b>	Svantek	<b>Serial Number:</b>	69558

---

**Calibration Procedure**

The sound level meter was allowed to stabilize for a suitable period, as described in the manufacturer's instruction manual, in laboratory conditions. The sound level meter was calibrated by carrying out the verification tests detailed in IEC 61672-3 (2006), Periodic tests, specification of sound level meters. Tolerances for verification procedures are specified in IEC 61672-1 (2003).

---

**Calibration Standards**

<b>Description</b>	<b>Serial Number</b>
National Instruments PXI-4461	19C91D2
Stanford Research DS360	123803

---

The standards used in this calibration are traceable to NIST and/or other National Measurement Institutes (NMI's) that are signatories of the International Committee of Weights and Measures (CIPM) mutual recognition agreement (MRA).

---

**Signed on behalf of Sonitus Systems:**



---

**Equipment Description**

**Model:** Svantek **Serial Number:** 69558  
**Model:** 977 **Microphone Model:** ACO 7052E

---

**Ambient Conditions**

Measurement conditions were within the tolerances defined in IEC 61672-1 and IEC 60942.

**Barometric Pressure:** 1030 hPa  
**Temperature:** 23.4 °C  
**Relative Humidity:** 39 %

---

**Results Summary**

IEC 61672 Test #	Test Description	Result
10	Self-generated noise	-
11	Frequency weighting (acoustical)	PASS
12	Frequency weighting (electrical)	PASS
13	Frequency and time weighting (1kHz)	PASS
14	Level linearity on reference level range	PASS
15	Level linearity with level range control	-
16	Toneburst response	PASS
17	Peak C sound level	PASS
18	Overload indication	PASS

---

As public evidence was available, from a testing organization responsible for approving the results of pattern evaluation tests, to demonstrate that the model of sound level meter fully conformed to the requirements for pattern evaluation described in IEC 61672:2003, the sound level meter tested is considered to conform to all the Class 1 requirements of IEC 61672:2003.

The manufacturer's guidelines concerning appropriate set up for measurement under various conditions should be observed during usage.

---

Prior to carrying out the verification tests the sound level meter was adjusted to read correctly using the acoustic calibrator held by the testing lab (Cirrus CR511ES, Serial number: 60871). The calibration procedure is described in the manufacturer's instruction manual.

Self-generated noise - IEC 61672-3 Test #10

SLM Measuring Mode: Leq

SLM Configuration	Freq. Weighting Network	SLM Reading
Microphone Installed	A	18.8
Microphone replaced by electrical input device fitted with short circuit	A	10.6
	C	10.6
	Z	10.6

Acoustical signal test of a frequency weighting - IEC 61672-3 Test #11

Range: reference level range

Frequency Weighting: C

Time Weighting: Slow

Input	Freq	Expected Level	Deviation	Tol +/-
94 dB	1000 Hz	94.0	0.0	1.0
	125 Hz	93.7	0.2	1.0
	4000 Hz	92.3	0.5	1.0

The frequency response was tested using an electrostatic actuator. Appropriate correction factors were applied where available from the manufacturer's instruction manual.

Electrical tests of frequency weighting - IEC 61672-3 Test #12

Range: reference level range

A-weighting

Freq	Expected Level	SLM Reading	Deviation	Tol +	Tol -
63	95.0	95.0	0.0	1.5	-1.5
125	95.0	94.9	-0.1	1.5	-1.5
250	95.0	94.9	-0.1	1.4	-1.4
500	95.0	95.0	0.0	1.4	-1.4
1000	95.0	95.0	0.0	1.1	-1.1
2000	95.0	95.0	0.0	1.6	-1.6
4000	95.0	95.1	0.1	1.6	-1.6
8000	95.0	95.1	0.1	2.1	-3.1
16000	95.0	94.7	-0.3	3.5	-17.0

C-weighting

Freq	Expected Level	SLM Reading	Deviation	Tol +	Tol -
63	95.0	95.0	0.0	1.5	-1.5
125	95.0	95.4	0.4	1.5	-1.5
250	95.0	95.0	0.0	1.4	-1.4
500	95.0	95.0	0.0	1.4	-1.4
1000	95.0	95.0	0.0	1.1	-1.1
2000	95.0	95.1	0.1	1.6	-1.6
4000	95.0	95.1	0.1	1.6	-1.6
8000	95.0	95.1	0.1	2.1	-3.1
16000	95.0	94.7	-0.3	3.5	-17.0

Linear

Freq	Expected Level	SLM Reading	Deviation	Tol +	Tol -
63	95.0	95.0	0.0	1.5	-1.5
125	95.0	95.0	0.0	1.5	-1.5
250	95.0	95.0	0.0	1.4	-1.4
500	95.0	95.0	0.0	1.4	-1.4
1000	95.0	95.0	0.0	1.1	-1.1
2000	95.0	95.0	0.0	1.6	-1.6
4000	95.0	95.0	0.0	1.6	-1.6
8000	95.0	95.0	0.0	2.1	-3.1
16000	95.0	95.0	0.0	3.5	-17.0

Frequency and Time Weightings at 1 kHz IEC 61672-3 Test #13

Range: reference level range

Time Weighting	Freq. Weighting	Expected Level	Deviation	Tol +/-
Fast	A	94.0	ref	
	C	94.0	0.0	0.2
Slow	A	94.0	0.0	0.2
LEQ	A	94.0	0.0	0.2

Linearity level on reference range - IEC 61672-3 Test #14  
 Input frequency: 8 kHz  
 SLM Measuring Mode: SPL

Range	Expected Level	SLM Reading	Deviation	Tol +/-
123 dB	94.0	94.0	0.0	1.1
	99.0	99.0	0.0	1.1
	104.0	104.0	0.0	1.1
	109.0	109.0	0.0	1.1
	114.0	114.0	0.0	1.1
	119.0	119.0	0.0	1.1
	124.0	124.0	0.0	1.1
	129.0	129.0	0.0	1.1
	134.0	134.0	0.0	1.1
	135.0	135.0	0.0	1.1
	136.0	136.0	0.0	1.1
	137.0	137.0	0.0	1.1
	89.0	89.0	0.0	1.1
	84.0	84.0	0.0	1.1
	79.0	79.0	0.0	1.1
	74.0	74.0	0.0	1.1
	69.0	69.0	0.0	1.1
	64.0	64.0	0.0	1.1
	59.0	59.0	0.0	1.1
	54.0	54.0	0.0	1.1
	49.0	49.1	0.1	1.1
	44.0	44.1	0.1	1.1
	39.0	39.1	0.1	1.1
	38.0	38.1	0.1	1.1
	37.0	37.2	0.2	1.1
	36.0	36.2	0.2	1.1
	35.0	35.3	0.3	1.1

Toneburst response - IEC 61672-3 Test #16

Range: reference level range

Burst Type	Response	Expected Level	SLM Reading	Deviation	Tol +	Tol -
0.25 ms	LAF <sub>MAX</sub>	111.0	110.8	-0.2	0.8	-0.8
2.0 ms	LAF <sub>MAX</sub>	120.0	119.9	-0.1	1.3	-1.3
200 ms	LAF <sub>MAX</sub>	137.0	137.0	0.0	1.3	-3.3
2.0 ms	LAS <sub>MAX</sub>	111.0	111.3	0.3	0.8	-0.8
200 ms	LAS <sub>MAX</sub>	130.6	130.6	0.0	1.3	-3.3

Peak C sound level - IEC 61672-3 Test #17

Range: reference level range

Pulse Type	Freq	Expected Level	SLM Reading	Deviation	Tol +/-
1 cycle	8 kHz	135.4	135.3	-0.1	2.4
Pos ½ cycle	500 Hz	137.4	137.4	0	1.4
Neg ½ cycle	500 Hz	137.4	137.4	0	1.4

Overload indication IEC 61672-3 Test #18

Test Description	Overload at	Meas. Diff. (Pos – Neg)	Tol +/-
Pos. ½ cycle at 4 kHz	142.5		
Neg. ½ cycle at 4 kHz	142.5		
Level difference		0.0	1.8

**Calibration Notes**

1. The manufacturer's instruction manual was accessed through the manufacturer's website.
2. The sound level meter was powered by a regulated 9V power supply provided by the testing laboratory.

ENVIRONMENTAL CONDITIONS

Temperature	Relative humidity	Ambient pressure
29 °C	37%	1005 hPa

TEST EQUIPMENT

Item	Manufacturer	Model	Serial no.	Description
1.	SVANTEK	SVAN 401	127	Signal generator
2.	SVANTEK	SVAN 912A	9537	Sound & Vibration Analyser
3.	KEITHLEY	2000	0910165	Digital multimeter
4.	SVANTEK	SV33	48876	Acoustic calibrator
5.	SVANTEK	ST02	-	Microphone equivalent electrical impedance (18pF)
6.	DYTRAN	1331A	1376	Reference accelerometer

CONFORMITY & TEST DECLARATION

1. Herewith Svantek company declares that this instrument has been calibrated and tested in compliance with the internal ISO9001 procedures and meets all specification given in the Manual(s) or respectively surpass them.
2. The acoustic calibration was performed using the Sound Calibrator and is traceable to the GUM (Central Office of Measures) reference standard - sound level calibrator type 4231 No 2292773.
3. The vibrational calibration was performed using the Back-to-Back Comparison method and is traceable to the GUM (Central Office of Measures) reference standard - accelerometer type 8305 No 1435233.
4. The information appearing on this sheet has been compiled specifically for this instrument. This form is produced with advanced equipment & procedures which permit comprehensive quality assurance verification of all data supplied herein.
5. This calibration sheet shall not be reproduced except in full, without written permission of the SVANTEK Ltd.

Calibration specialist: Krzysztof Kubeł



Test date: 2018-06-01



ISO9001 certified

FACTORY CALIBRATION DATA OF THE SVAN 977A No. 69556

with preamplifier SVANTEK type SV12L No. 72145 and microphone ACO PACIFIC type 7052E No. 69608

SOUND LEVEL METER

1. CALIBRATION (electrical)

LEVEL METER function; Characteristic: A;  $f_{ref}=1$  kHz; Input signal =110.9 dB;

Range	Low (120dB)	High (137dB)
Indication [dB]	114.0	114.0
Error [dB]	0.0	0.0

2. CALIBRATION\* (acoustical)

LEVEL METER function; Range: High; Reference frequency: 1000 Hz; Sound Pressure Level: : 113.99 dB

Characteristic	Correct value [dB]	Indication [dB]	Error [dB]
Z	113.99	114.02	0.03
A	113.99	114.02	0.03
C	113.99	114.02	0.03

Calibration measured with the microphone ACO PACIFIC type 7052E No. 69608. Calibration factor: 0.63 dB.

3. LINEARITY TEST\* (electrical)

LEVEL METER function; Range: Low; Characteristic: A;  $f_{ref}=31.5$  Hz

Nominal result LEQ [dB]	24.0	25.0	26.0	28.0	30.0	40.0	60.0	80.0
Error [dB]	0.1	0.1	0.1	0.0	0.0	0.0	0.0	0.0

LEVEL METER function; Range: Low; Characteristic: A;  $f_{ref}=1000$  Hz

Nominal result LEQ [dB]	24.0	25.0	26.0	28.0	30.0	40.0	60.0	80.0	100.0	120.0
Error [dB]	0.1	0.1	0.0	0.0	-0.0	-0.0	-0.0	-0.0	-0.0	-0.0

LEVEL METER function; Range: Low; Characteristic: A;  $f_{ref}=8000$  Hz

Nominal result LEQ [dB]	24.0	25.0	26.0	28.0	30.0	40.0	60.0	80.0	100.0	119.0
Error [dB]	0.1	0.0	0.0	-0.0	-0.0	-0.0	-0.0	-0.0	0.0	-0.0

LEVEL METER function; Range: High; Characteristic: A;  $f_{ref}=31.5$  Hz

Nominal result LEQ [dB]	35.0	36.0	37.0	38.0	40.0	60.0	80.0	97.0
Error [dB]	0.1	0.1	0.0	0.0	0.0	-0.0	-0.0	0.0

LEVEL METER function; Range: High; Characteristic: A;  $f_{ref}=1000$  Hz

Nominal result LEQ [dB]	35.0	36.0	37.0	38.0	40.0	60.0	80.0	100.0	120.0	137.0
Error [dB]	0.1	0.1	0.0	0.0	0.0	-0.0	-0.0	-0.0	-0.0	0.0

LEVEL METER function; Range: High; Characteristic: A;  $f_{ref}=8000$  Hz

Nominal result LEQ [dB]	35.0	36.0	37.0	38.0	40.0	60.0	80.0	100.0	120.0	136.0
Error [dB]	0.1	0.1	0.0	0.1	0.0	-0.0	-0.0	0.0	-0.0	0.0

4. TONE BURST RESPONSE\*

LEVEL METER function; Characteristic: A;  $f_{ref}=4000$  Hz; Burst duration: 2 s

Range: Low; Steady level nominal result = 117dB

Result	Detector	Duration [ms]	1000	500	200	100	50	20	10	5	2	1	0.5	0.25
MAX	Fast	Indication [dB]	117.0	116.9	116.0	114.4	113.3	108.7	105.8	102.9	99.0	96.0	92.9	89.9
		Error [dB]	0.0	0.0	0.0	-0.0	-0.0	-5.1	0.0	-0.0	-0.1	-0.1	-0.1	-0.1
	Slow	Indication [dB]	115.0	113.0	106.6	106.8	103.9	100.0	97.0	94.0	90.0	-	-	-
		Error [dB]	0.0	0.0	-0.0	-0.0	-0.0	-0.0	-0.0	-0.0	-0.0	-	-	-
SEL	-	Indication [dB]	117.0	114.0	110.9	107.0	104.0	100.0	97.0	94.0	90.0	87.0	83.9	80.9
		Error [dB]	0.0	-0.0	0.0	0.0	-0.0	0.0	-0.0	-0.0	-0.0	-0.1	-0.1	-0.1